

#### Product Description for 3M 1860 N95 Mask

#### Product #EK-FM-N1

Features:

- Bacterial filtration efficiency (BFE) ≥ 99%
- NIOSH approved
- FDA cleared for use as surgical mask
- Collapse resistant cup shape design
- Braided head strap
- Adjustable nose clip and cushioning nose foam
- One size fits most
- No exhalation valve
- Contains no natural rubber latex
- Helps protect against certain airborne biological particles
- Fluid resistant and disposable
- Advanced electrostatic media
- Flammability Class 1
- Fluid resistance: 120mm Hg
- Color: teal
- Shelf life: 5 years



For pricing and product information, we can be reached as follows: Clean and Healthy Products, Inc. info@cleanandhealthyproducts.com Phone: (818) 903-2790



# FDA Listing for 3M 1860 N95 Mask

#### Product #EK-FM-N1

FDA U.S. FOOD & DRUG						
Home Food Dru	gs Medical Devices Radiation-Emi	tting Products	Vaccines, Blood & Biologics	Animal & Veterinary	Cosmetics	Tobacco Products
	dical Devices O Databases					i 🖬 🔛
CDRH SuperSearch			arse Events   Recalls   PMA   HDE   s   X-Ray Assembler   Medsun Reports			
	New Search			Back To Se	earch Results	_
	Device Classification Name 510(K) Number Device Name Applicant	K955382	360 HEALTH CARE PARTICULAT 7 51-3 AND 270	E RESPIRATOR TYPE N	195	
	Applicant Contact Correspondent	Dorothea Von 3M COMPANY BUILDINGS 2 St. Paul, MN	/ 51-3 AND 270			
	Correspondent Contact Regulation Number Classification Product Code Date Received Decision Date	11/24/1995 05/01/1996				
	Decision Regulation Medical Specialt 510k Review Panel Summary Type	y General & Plas General Hospi <u>Summary</u> Traditional	• •			
	Reviewed By Third Party Combination Product	No No				

For pricing and product information, we can be reached as follows:

Clean and Healthy Products, Inc.

info@cleanandhealthyproducts.com

Phone: (818) 903-2790



## NIOSH Listing for 3M 1860 N95 Mask

### Product #EK-FM-N1

National Persona	Protective Technology Laboratory	(NPPTL)
ed Equipment List	Certified Equipment List	
neu equipment else		Promoting productive workplaces through safety and health research
al Cautions and 🛛 🕂		
tions	Approval Details	
tions of Terms	Schedule	84A
Manufacturers Names	Approval Number	0006
PTL Homepage	Manufacturer	<u>3M Company</u>
Z Index	Approval Date	8/17/1995
Respirator Users	Product Line	1860 and 1860S
Respirator Manufacturers	For Protection Against	N95
tective Clothing and	Service Time (minutes)	
Ensembles	Facepiece Type	Filtering Facepiece
tective Technology gram at NIOSH	Is the respirator powered?	Non-powered
pirator Trusted-Source	Number of Filters	1
rmation	Is the the filter replaceable?	No
roved Particulate	Filter Location	Filtering Facepiece
Filtering Facepiece Respirators	Cartridge or Canister	
Certified Equipment List	Number of cartridge or canisters	
CEL)	Location of cartridge or canisters	
pirator User Notices	Is the cartridge or canister replacable?	No
itact NPPTL	Self-contained breathing apparatus (SCBA) Type	
	Supplied air respirator (SAR) Type	
	Self-contained breathing apparatus (SCBA) Use	
	Airflow	
	Breathing Gas	
	Percentage of Oxygen in Breathing Gas	
	Regulator Location	
	Low Temperature (Celsius)	
	Low Temperature (Fahrenheit)	
	Exhalation Valve?	No
	Self Rescuer?	No
	Is the approval obsolete?	No
	Private Label	
	Percentage of Oxygen in Breathing Gas	
	Regulator Location	
	Low Temperature (Celsius)	
	Low Temperature (Fahrenheit)	
	Exhalation Valve?	No
	Self Rescuer?	No
	Is the approval obsolete?	No
	Private Label	
	Model	
	Rated Duration	
	Cylinder pounds per square inch gauge (PSIG)	
	Chemical, biological, radiological and nuclear (CBRN)	
	Approval Date	

For pricing and product information, we can be reached as follows:

Clean and Healthy Products, Inc.

info@cleanandhealthyproducts.com

Phone: (818) 903-2790